

Anatek Labs, Inc.

Anatek Moscow - 1282 Alturas Drive - Moscow, ID 83843 - 208-883-2839 - moscow@anateklabs.com - FL NELAP E87893
Anatek Spokane - 504 E Sprague Ste. D - Spokane, WA 99202 - 509-838-3999 - spokane@anateklabs.com - FL NELAP E871099
Anatek Yakima - 4802 Tieton Drive - Yakima, WA 98908 - 509-225-9404 - yakima@anateklabs.com - FL NELAP E871190
Anatek Wenatchee - 3019 Gs Center Rd - Wenatchee, WA 98801 - 509-701-8362

Client: Pelican Point
Address: PO Box 3388
Coeur d Alene, ID 83816
Attn: Leslie Rayner

Work Order: WFA0402
Project: Nitrate
Reported: 1/15/2025 17:31

Analytical Results Report


System ID# **66800** System Name: **Pelican Point**
Reference Number: **WFA0402-01** Collect Date: **01/13/25 09:50** County: **Grant**
DOH Source #: **02** Sample Type: **Before** Multiple Source Nos:
Matrix: **Drinking Water** Composition: **S - Single source samples**
Date Received: **01/13/25 15:55** Sample Purpose: **RC - Routine/Compliance Sample**
Sample Location: **Well Head**

Lab/Sample Number: 112-40201

Inorganics

DOH #	Analyte	Result	Units	LRL	SDRL	Trigger	MCL	Analyzed	Analyst	Method	Qualifier
0020	Nitrate as N	1.63	mg/L	0.100	0.5	5	10	1/13/25 16:43	BAM	EPA 300.0	

Authorized Signature,



Brock Gerger for Kathleen Sattler, Lab Manager

LRL Lab Reporting Limit
SDRL State Detection Reporting Limit
ND Not Detected
MCL EPA's Maximum Contaminant Level
Dry Sample results reported on a dry weight basis
SAL State Action Level
* Not a certified analyte

This report shall not be reproduced except in full, without the written approval of the laboratory
The results reported related only to the samples indicated.

Anatek Labs, Inc.

Anatek Moscow - 1282 Alturas Drive - Moscow, ID 83843 - 208-883-2839 - moscow@anateklabs.com - FL NELAP E87893
Anatek Spokane - 504 E Sprague Ste. D - Spokane, WA 99202 - 509-838-3999 - spokane@anateklabs.com - FL NELAP E871099
Anatek Yakima - 4802 Tieton Drive - Yakima, WA 98908 - 509-225-9404 - yakima@anateklabs.com - FL NELAP E871190
Anatek Wenatchee - 3019 Gs Center Rd - Wenatchee, WA 98801 - 509-701-8362

Client: Pelican Point
Address: PO Box 3388
Coeur d Alene, ID 83816
Attn: Leslie Rayner

Work Order: WFA0402
Project: Nitrate
Reported: 1/15/2025 17:31

Certifications

Code	Description	Facility	Number
W WA DOE	Washington Department of Ecology	Anatek-Spokane, WA	C585

Anatek Labs, Inc.

Anatek Moscow - 1282 Alturas Drive - Moscow, ID 83843 - 208-883-2839 - moscow@anateklabs.com - FL NELAP E87893
Anatek Spokane - 504 E Sprague Ste. D - Spokane, WA 99202 - 509-838-3999 - spokane@anateklabs.com - FL NELAP E871099
Anatek Yakima - 4802 Tieton Drive - Yakima, WA 98908 - 509-225-9404 - yakima@anateklabs.com - FL NELAP E871190
Anatek Wenatchee - 3019 Gs Center Rd - Wenatchee, WA 98801 - 509-701-8362

Client: Pelican Point
Address: PO Box 3388
Coeur d Alene, ID 83816
Attn: Leslie Rayner

Work Order: WFA0402
Project: Nitrate
Reported: 1/15/2025 17:31

Quality Control Data

Inorganics

Analyte	Result	Qual	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit
---------	--------	------	--------------------	-------	----------------	------------------	------	----------------	-----	--------------

Batch: BFA0409 - W Ions

Blank (BFA0409-BLK1)

Nitrate/N ND 0.100 mg/L

Prepared & Analyzed: 01/13/25 10:27

LCS (BFA0409-BS1)

Nitrate/N 3.98 mg/L

Prepared & Analyzed: 01/13/25 11:09

4.00 99.6 90-110

MRL Check (BFA0409-MRL1)

Nitrate/N 0.131 mg/L

Prepared & Analyzed: 01/13/25 11:30

0.100 131 50-150

Matrix Spike (BFA0409-MS1)

Nitrate/N 4.56 mg/L

Source: WFA0393-01

Prepared & Analyzed: 01/13/25 12:11

4.00 0.0860 112 80-120

Matrix Spike Dup (BFA0409-MSD1)

Nitrate/N 4.30 mg/L

Source: WFA0393-01

Prepared & Analyzed: 01/13/25 12:32

4.00 0.0860 105 80-120 5.75 20